Contribution ID: 153

Type: Contributed Oral

# Design of a cryogenic capacitive liquid level sensor

Sunday, 18 February 2024 10:30 (15 minutes)

The Light-Only Liquid Xenon (LoLX) experiment is designed to study scintillation light and characterize the performance of silicon photomultipliers (SiPMs) in a liquid xenon (LXe) medium. To ensure the SiPMs are functioning properly, it is necessary to verify that all of them are submerged. LoLX therefore requires a LXe level sensor. Outgassing requirements, cryogenic temperature, space constraints and the low relative permittivity of LXe ( $\epsilon_{LXe} = 1.85$ ) prohibit the use of standard liquid level sensors.

We developed a highly sensitive capacitive liquid level sensor which consists of sinusoidal electrodes inside a flexible Kapton printed circuit board. The liquid level sensor is 30 cm long with a capacitance of  $35.4 \pm 0.1$  pF. The capacitance is read by a FDC1004 evaluation board and an Arduino Uno, and then sent to the LoLX slow control system to be converted to a liquid level. From measurements made in sunflower oil (\$\epsilon\_{oil} = 3.29}), we calculated that the sensitivity of the sensor in LXe will be  $0.1889 \pm 0.0001$  pF/cm, which is orders of magnitude higher than liquid level sensors used by other groups. We plan to cross-calibrate the sensor in LXe with temperature-sensitive resistors as fixed-point sensors once it has been installed. Commissioning in LXe is planned at the next run of LoLX in March 2024.

## Your Email

sandhya.rottoo@mail.mcgill.ca

### Supervisor

Thomas Brunner

#### **Supervisor Email**

sandhya.rottoo@mail.mcgill.ca

### Affiliation

McGill University

## Your current academic level,

Undergraduate student

#### Primary author: ROTTOO, Sandhya (McGill University)

Presenter: ROTTOO, Sandhya (McGill University)

Session Classification: Morning 6 - Feb. 18, 2024

Track Classification: Physics Beyond the Standard Model